

## EC - DECLARATION OF CONFORMITY



No.: IH60\_CE\_DoC, valid to 2019-07-02  
Manufacturer: Beurer GmbH  
Address: Söflinger Str. 218  
89077 Ulm, Germany

Förderndes Mitglied der  
Deutschen Hochdruckliga  
Förderndes Mitglied der  
Deutschen Diabetes Stiftung

This declaration of conformity is issued under our sole responsibility.

UMDNS code and name: 12-128 Nebulizer, Aerosol  
Category code and category: 04 Electrical and mechanical medical devices  
Classification Rule/ Class: Rule 11/ Class IIa  
Product name: IH60  
Expiry Date: 2019-07-02

The object of the declaration described above is in conformity with the relevant Community harmonisation legislation:

Directives\_ Date of last revision: CD 93/42/EEC (MDD) amended with 2007/47/EEC\_2010-03-21  
CD 2002/96/EC (WEEE)\_2003-01-27  
Reg. No.: DE 96226384  
CD 2011/65/EC (ROHS)\_2011-06-08

Conformity assessment procedure: Annex VII (Internal Production Control) in conjunction with Annex V (QA Production) of CD 93/42/EEC (MDD)

Relevant harmonised standards: IEC 60601-1: 2012  
EN 60601-1-2: 2015  
EN 13544-1: 2007 + A1: 2009

The notified body mdc - medical device certification GmbH, located at Kriegerstr. 6, D-70191 Stuttgart, identification number 0483, issued in the course of conformity assessment procedure the following certificate:

Certificate No.\_issued on: D1311700024\_2015-12-08, valid to 2019-07-02  
Signed for and on behalf of: Beurer GmbH  
Place, date of issue: Ulm, 2019-03-25  
Name, function and signature: Werner Meternek, Director R&D

*W. Meternek*  
Beurer GmbH  
Söflinger Straße 218 • 89077 Ulm