EC DECLARATION OF CONFORMITY



NRS Healthcare, Blaris Industrial Estate 4C Altona Road, Lisburn, Northern Ireland BT27 5QB (SRN: XI-IM-000004492)

Declare, upon our sole responsibility, that our product;

## Item Number (UDI/barcode): P23499 (GTIN 5038948234992) Item name: Swedish Bath Rail



## Being a Class I Medical Device (GMDN code: 35584)

Complies with the appropriate requirements of EC Council Directives:-

Regulation 2017/745 - Medical Devices

The product complies with the appropriate requirements of the listed harmonised or national standards:

BS EN 12182:2012 Assistive products for persons with disability. General requirements & test methods

Intended Use:

The Swedish Bath Rail is designed to help steady the user while transferring into or out of the bath. It can be fitted to most standard baths.

Jan Shirt

tan Smith Quality & Compliance Illanager

Date: January 2021

CE