

EC DECLARATION OF CONFORMITY



NRS Healthcare , Sherwood House, Cartwright Way, Forest Business Park,
Bardon Hill, Coalville, Leicestershire, LE67 1UB

Declare, upon our sole responsibility, that our product;

Item Number: H87139, L64348, L29324, L29312
Item name: Tripod & Tetrapod Walking Sticks

Complies with the appropriate requirements of EC Council Directives:-


93/42/EEC as amended by Directive 2007/47/EC – Medical Devices
94/62/EC and amendments – Packaging & Packaging Waste Directive

The product complies with the appropriate requirements of the listed harmonised or national standards:

EN1985:1998 Walking aids – general requirements & test methods

Intended Use:

Tripods and Tetrapods are walking aids with 3 or more legs designed to give the user a more stable support than with a single stick.


Ian Smith
Quality & Compliance Manager

Date: 18/3/19

CE