

EC DECLARATION OF CONFORMITY



NRS Healthcare , Barratt House, Cartwright Way, Forest Business Park,
Bardon Hill, Coalville, Leicestershire, LE67 1UB

Declare, upon our sole responsibility, that our product;

Item Number: M39634

Item name: Mobility Care 4 Wheel Aluminium Rollator

Complies with the appropriate requirements of EC Council Directives:-

93/42/EEC as amended by Directive 2007/47/EC – Medical Devices
94/62/EC and amendments – Packaging & Packaging Waste Directive
Furniture & Furnishings (Fire) (Safety) Regulations 1988

The product complies with the appropriate requirements of the listed harmonised or national standards:

BS EN ISO 11199:2005	Walking aids manipulated by both arms – requirements and test methods: Part 2 Rollators.
BS5852:2006	Methods of test for assessment of the ignitability of upholstered seating by smouldering and flaming ignition sources

Intended Use:

the Mobility Care Aluminium Rollator is designed to provide support and assistance to those unable to, or lack the confidence to, walk without some level of support


Ian Smith
Quality & Compliance Manager
Date: 16/3/16

