

EC DECLARATION OF CONFORMITY



NRS Healthcare , Barratt House, Cartwright Way, Forest Business Park,
Bardon Hill, Coalville, Leicestershire, LE67 1UB

Declare, upon our sole responsibility, that our product;

Item Number: M07232, M07244, M07256, M07268, M07279
Item name: NRS Deep Back Support Sling

Complies with the appropriate requirements of EC Council Directives:-

93/42/EEC as amended by Directive 2007/47/EC – Medical Devices

The product complies with the appropriate requirements of the listed harmonised or national standards:

BS EN 12182:2012	Assistive products for persons with disability. General requirements & test methods
EN10535:2006	Hoists for the transfer of disabled persons – requirements & test methods
SHTM87	Scottish Health Technical Memorandum No 87 - Textiles & Furniture

Intended Use: The Deep Back Support Slings range are general purpose slings designed to be used with a patient hoist to lift and transfer a person who is unable to transfer independently and must therefore be hoisted. The sling is designed for use with most hoists with "coat hanger" style spreader bars.


John Turner
New Product Development Manager

Date: 29 September 2014

