

# EC DECLARATION OF CONFORMITY



NRS Healthcare , Barratt House, Cartwright Way, Forest Business Park,  
Bardon Hill, Coalville, Leicestershire, LE67 1UB

*Declare, upon our sole responsibility, that our product;*

**Item Number: F19739**

**Item name: Malvern Vinyl Seat Perching Stool - standard**

*Complies with the appropriate requirements of EC Council Directives:-*

93/42/EEC as amended by Directive 2007/47/EC – Medical Devices  
94/62/EC and amendments – Packaging & Packaging Waste Directive

*The product complies with the appropriate requirements of the listed harmonised or national standards:*

BS EN 12182:2012	Assistive products for persons with disability. General requirements & test methods
BS6807:2006	Methods of test for assessment of ignitability of mattresses, upholstered divans & upholstered bed bases with flaming types of primary & secondary sources of ignition

**Intended Use:**

For users to temporarily 'sit' when performing tasks where standing for a long time may be difficult.

  
Ian Smith  
Quality & Compliance Manager

Date:11/4/16

